

**The Security Printing Corporation (Bangladesh) Ltd. (SPCBL),
Gazipur.**

Minutes of pre-bid meeting regarding Local & International e-tender nos. 11 & 12/2018-2019 for supply, installation, testing & commissioning of a Spectrophotometer.

A pre-bid meeting on the captioned e-tenders was held on the 10th October, 2018 at the Board Room of the SPCBL. The Managing Director of the SPCBL presided over the meeting. The other representatives of the SPCBL who attended the meeting were as follows :

1. Mr. Md. Kalimulla, General Manager (Foreign Purchase), SPCBL, Gazipur.
2. Mr. Md. Anwar Hossain Chowdhury, General Manager (OSP), SPCBL, Gazipur.
3. Mr. Ehteshamul Karim, General Manager (Finance and Accounts), SPCBL, Gazipur.
4. Mr. Md. Khasruzzaman, Deputy General Manager (O. R & QC), SPCBL, Gazipur.
5. Mr. Md. Jamil Akhtar Shahjada, Additional Chief Engineer & in-charge, Chief Engineer, SPCBL, Gazipur.
6. Mr. Md. Humayun Kabir, Executive Engineer (Electrical), SPCBL, Gazipur.
7. Mr. A. K. M. Iqbal Hossain, Deputy Manager (Foreign Purchase), SPCBL, Gazipur.
8. Mr. Mostafa Selim Durrani, Deputy Manager (R & QC), SPCBL, Gazipur.

On the other hand the following representatives of the intending tenderers were present at the meeting :

1. Mr. Faisal Delwar, Director, Modern Technology, Dhaka.
2. Mr. Manik Sarker, Application Manager, Modern Technology, Dhaka.
3. Mr. Md. Abdus Salam, CEO, Alltech Service, Dhaka.

The President welcomed the representatives of the intending tenderers and invited them to raise any issue they feel to be explicated on the technical specifications and the terms and conditions set forth in the concerned e-tender schedule. The SPCBL side informed the meeting that the technical specifications mentioned in the concerned tender schedule has been prepared as per the existing Spectrophotometer of the SPCBL procured almost two decades back. In this situation, the participants from both sides expressed their opinions/ideas regarding the tendered technical specifications of the spectrophotometer as well as the latest technologies/inventions in this regard. In this context, it has been observed through mutual open discussion that meanwhile technology has developed to a far extend and technical specifications of the tender in question is not updated which should be amended in some parameters accordingly so that the Spectrophotometer with the latest technology can be procured. However, after a long and fruitful discussion, the following amendments of the e-tender schedules have been adopted :

| Tender Clauses | Amendments of e-Tender Schedule adopted in the pre-bid meeting. |
|-------------------------|---|
| D.9 (Spectral data) : | Reported at 10 nanometer intervals across the visible spectrum 360 nm to 780 nm instead of existing. |
| D. 10 (Aperture Size) : | LAV 25 mm, 17, SAV 10/11mm, USAV 6 mm & USAV 0.020 mm (0.008") instead of existing. |
| D. 19 (Provision) : | d) Output of reflectance values in standard intervals which shall be shown in computer and printer (included) instead of existing. f) Manual/Automatic input of spectral data instead of existing. |

Having no other issues for discussion, the chair concluded the meeting with thanks to all. This minutes will be treated as an integral part of the e-tender schedule.

(SK. Azizul Haque),
Managing Director, SPCBL.